
Midmark M11 - v1.0.0
Total Cycles: 2790

Name: RITTER 1
Sterilizer ID: V2178198
01/18/2023 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.3 F 0.0
2:00 208.7 F 2.4
4:00 208.0 F 7.1
6:00 217.1 F 12.8
8:00 225.8 F 7.7
10:00 238.9 F 13.1
12:00 239.2 F 10.7
14:00 237.0 F 9.1
16:00 251.3 F 16.1
18:00 263.4 F 23.1
20:00 271.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.6 F 29.1
2:00 271.6 F 28.5
3:00 272.7 F 29.2
4:00 272.0 F 28.6

Min 271.6 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.5 F 0.0
5:00 234.8 F 0.0
10:00 230.2 F 0.0
15:00 228.7 F 0.0
20:00 229.1 F 0.0
25:00 229.1 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

01/18/2023 10:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____