

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2557

Name: RITTER2  
Sterilizer ID: V2313228  
06/23/2023 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
2:00 186.4 F 3.4  
4:00 195.8 F 8.5  
6:00 221.4 F 9.3  
8:00 229.6 F 9.7  
10:00 234.2 F 8.8  
12:00 244.1 F 13.0  
14:00 244.4 F 12.8  
16:00 257.4 F 19.7  
18:00 269.1 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.7 F 29.7  
2:00 272.1 F 29.0  
3:00 271.4 F 29.0  
4:00 272.6 F 29.8  
  
Min 271.3 F 28.7  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 210.8 F 0.0  
10:00 219.0 F 0.0  
15:00 220.2 F 0.0  
20:00 219.7 F 0.0  
25:00 218.9 F 0.0  
30:00 217.4 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 19:08  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:11  
CYCLE COMPLETE

06/23/2023 11:11 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_