
Midmark M11 - v1.0.0
Total Cycles: 3568

Name: RITTER 1
Sterilizer ID: V2178198
07/24/2023 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.6 F 0.0
2:00 170.6 F 0.5
4:00 182.0 F 3.4
6:00 200.3 F 7.0
8:00 215.1 F 11.2
10:00 218.0 F 5.1
12:00 231.0 F 9.4
14:00 236.5 F 10.6
16:00 241.3 F 11.5
18:00 239.1 F 9.9
20:00 251.9 F 16.0
22:00 262.6 F 22.2
24:00 272.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.5 F 28.7
2:00 271.6 F 28.0
3:00 271.7 F 28.5
4:00 272.3 F 28.7

Min 271.1 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 214.2 F 0.0
10:00 204.0 F 0.0
15:00 198.4 F 0.0
20:00 195.1 F 0.0
25:00 194.4 F 0.0
30:00 205.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:27
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:02:58
CYCLE COMPLETE

07/24/2023 12:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____