
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001701
13:25 07/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:19
STERILIZING 3:21
136.0 C 222kPa 3:21
136.2 C 224kPa 3:51
136.2 C 224kPa 4:21
136.2 C 225kPa 4:51
136.4 C 223kPa 5:21
136.4 C 223kPa 5:51
136.4 C 224kPa 6:21
136.4 C 223kPa 6:51
136.4 C 223kPa 6:51
VENTING 6:51
AIR DRYING 7:43
CYCLE COMPLETE 8:05

Digital Signature #
FB700433328C9844
