
Midmark M11 - v1.0.0
Total Cycles: 1686

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2022 12:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.0 F 0.0
2:00 183.6 F 0.2
4:00 199.0 F 2.8
6:00 214.6 F 7.2
8:00 229.3 F 12.0
10:00 235.2 F 8.7
12:00 240.4 F 10.3
14:00 246.6 F 13.4
16:00 245.3 F 12.7
18:00 258.9 F 19.9
20:00 270.7 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.8 F 28.7
2:00 271.7 F 28.0
3:00 272.1 F 28.6
4:00 272.6 F 28.6

Min 271.3 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 226.3 F 0.0
10:00 232.9 F 0.0
15:00 218.4 F 0.0
20:00 217.9 F 0.0
25:00 208.4 F 0.0
30:00 215.3 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 20:48
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:01
CYCLE COMPLETE

03/01/2022 01:55 PM
ACCEPTED: X
REJECTED:
Operator: U02