
Midmark M11 - v1.0.0
Total Cycles: 3300

Name: RITTER 1
Sterilizer ID: V2178198
05/29/2023 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.9 F 0.0
2:00 186.4 F 0.0
4:00 191.1 F 4.2
6:00 202.8 F 8.4
8:00 239.5 F 13.0
10:00 236.7 F 8.9
12:00 237.7 F 10.5
14:00 243.9 F 12.0
16:00 240.8 F 10.4
18:00 253.1 F 16.8
20:00 264.2 F 23.2
22:00 271.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.1 F 29.1
2:00 272.1 F 29.2
3:00 272.1 F 29.2
4:00 272.1 F 29.2

Min 271.8 F 28.7
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.5 F 0.0
5:00 223.3 F 0.0
10:00 219.2 F 0.0
15:00 230.3 F 0.0
20:00 228.6 F 0.0
25:00 227.1 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 22:44
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:11
CYCLE COMPLETE

05/29/2023 04:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____