
Midmark M11 - v1.0.0
Total Cycles: 4510

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2024 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.4 F 0.0
2:00 114.5 F 0.7
4:00 120.6 F 3.0
6:00 148.4 F 6.0
8:00 173.3 F 9.4
10:00 193.4 F 13.2
12:00 210.6 F 5.1
14:00 219.4 F 8.8
16:00 231.7 F 13.0
18:00 233.5 F 9.2
20:00 241.0 F 11.0
22:00 243.6 F 12.5
24:00 255.0 F 18.3
26:00 265.3 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.3 F 29.3
2:00 271.5 F 28.7
3:00 271.7 F 29.2
4:00 272.2 F 29.2

Min 271.1 F 28.5
Max 272.3 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.8 F 0.0
5:00 201.4 F 0.0
10:00 197.8 F 0.0
15:00 217.6 F 0.0
20:00 239.9 F 0.0
25:00 228.0 F 0.0
30:00 221.0 F 0.0

Drying Complete

FILLING: 0:57
HEATING: 27:56
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:06:41
CYCLE COMPLETE

05/01/2024 02:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____