
Midmark M11 - v1.0.5
Total Cycles: 558

Name: RITTER2
Sterilizer ID: V2313228
11/17/2021 12:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.6 F 0.0
2:00 176.7 F 3.8
4:00 199.2 F 9.2
6:00 216.0 F 7.1
8:00 230.2 F 11.0
10:00 233.1 F 9.0
12:00 238.2 F 9.7
14:00 248.4 F 15.3
16:00 261.1 F 22.4
18:00 272.5 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.8 F 29.2
2:00 272.2 F 29.5
3:00 272.3 F 29.8
4:00 271.7 F 29.7

Min 271.5 F 29.2
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 205.4 F 0.0
10:00 205.6 F 0.0
15:00 206.0 F 0.0
20:00 206.2 F 0.0
25:00 205.3 F 0.0
30:00 205.4 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:40
CYCLE COMPLETE

11/17/2021 01:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____