
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000075
12:30 07/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:09
STERILIZING 4:33
136.4 C 224kPa 4:33
136.2 C 224kPa 5:03
136.3 C 224kPa 5:33
136.3 C 222kPa 6:03
136.2 C 223kPa 6:33
136.1 C 222kPa 7:03
136.0 C 221kPa 7:33
136.2 C 224kPa 8:03
136.2 C 224kPa 8:03
VENTING 8:03
AIR DRYING 8:55
CYCLE COMPLETE 23:55

Digital Signature #
F7C23A332DBEFBDF
