

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4164

Name: RITTER 1  
Sterilizer ID: V2178198  
12/28/2023 12:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.0 F 0.0  
2:00 143.2 F 1.9  
4:00 152.8 F 5.0  
6:00 210.4 F 9.4  
8:00 211.8 F 5.9  
10:00 225.1 F 7.2  
12:00 233.9 F 11.7  
14:00 233.4 F 7.8  
16:00 244.7 F 13.2  
18:00 242.0 F 11.3  
20:00 254.2 F 17.5  
22:00 265.2 F 24.0  
24:00 271.8 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.5 F 28.7  
2:00 272.7 F 29.1  
3:00 271.5 F 28.7  
4:00 272.5 F 28.7

Min 271.4 F 28.4  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.6 F 0.0  
5:00 246.9 F 0.0  
10:00 229.7 F 0.0  
15:00 225.0 F 0.0  
20:00 230.6 F 0.0  
25:00 229.8 F 0.0  
30:00 217.6 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 24:30  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:03:05  
CYCLE COMPLETE

12/28/2023 01:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_