

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4120

Name: RITTER2  
Sterilizer ID: V2313228  
11/14/2024 03:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.7 F 0.0  
2:00 138.6 F 1.5  
4:00 153.2 F 4.5  
6:00 181.1 F 8.3  
8:00 202.7 F 12.0  
10:00 213.3 F 6.1  
12:00 225.3 F 10.6  
14:00 229.8 F 7.7  
16:00 241.3 F 12.7  
18:00 240.5 F 11.2  
20:00 252.3 F 17.3  
22:00 262.9 F 23.7  
24:00 272.4 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 30.0  
1:00 272.4 F 30.4  
2:00 272.3 F 30.2  
3:00 272.3 F 30.1  
4:00 272.3 F 30.0

Min 272.1 F 29.9  
Max 272.5 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
5:00 220.9 F 0.0  
10:00 198.9 F 0.0  
15:00 190.6 F 0.0  
20:00 191.0 F 0.0  
25:00 189.1 F 0.0  
30:00 190.6 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 24:57  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:03:43  
CYCLE COMPLETE

11/14/2024 04:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_