
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001307
17:32 19/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:50
STERILIZING 4:58
136.4 C 224kPa 4:58
136.3 C 224kPa 5:28
136.4 C 225kPa 5:58
136.4 C 225kPa 6:28
136.3 C 223kPa 6:58
136.3 C 223kPa 7:28
136.3 C 224kPa 7:58
136.1 C 223kPa 8:28
136.1 C 223kPa 8:28
VENTING 8:28
AIR DRYING 9:24
CYCLE COMPLETE 14:50

Digital Signature #
68E78550EB5D98BE
