

-----  
Midmark M11 - v1.0.5  
Total Cycles: 262

Name: RITTER2  
Sterilizer ID: V2313228  
08/04/2021 10:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.9 F 0.0  
2:00 131.9 F 1.3  
4:00 150.1 F 4.1  
6:00 172.8 F 7.7  
8:00 193.3 F 12.1  
10:00 212.1 F 5.2  
12:00 222.4 F 9.4  
14:00 232.2 F 11.0  
16:00 237.7 F 11.0  
18:00 235.8 F 9.1  
20:00 248.7 F 15.1  
22:00 259.0 F 21.0  
24:00 268.8 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.4 F 29.6  
2:00 271.3 F 29.1  
3:00 272.3 F 29.7  
4:00 271.8 F 29.1

Min 271.3 F 29.0  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.0 F 0.0  
5:00 226.6 F 0.0  
10:00 216.1 F 0.0  
15:00 216.5 F 0.0  
20:00 199.0 F 0.0  
25:00 199.0 F 0.0  
30:00 185.2 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 25:14  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:03:36  
CYCLE COMPLETE

08/04/2021 11:55 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01