
Midmark M11 - v1.0.0
Total Cycles: 4336

Name: RITTER 1
Sterilizer ID: V2178198
02/27/2024 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 74.3 F 0.2
4:00 74.8 F 1.0
6:00 84.2 F 2.9
8:00 104.3 F 5.2
10:00 134.0 F 7.8
12:00 160.1 F 10.8
14:00 208.2 F 6.7
16:00 210.0 F 4.3
18:00 220.4 F 7.3
20:00 229.3 F 10.6
22:00 233.1 F 8.6
24:00 237.5 F 10.5
26:00 235.9 F 8.8
28:00 245.6 F 13.2
30:00 255.7 F 18.5
32:00 265.1 F 24.2
34:00 271.2 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.4 F 29.1
2:00 272.4 F 29.5
3:00 271.5 F 29.0
4:00 272.3 F 29.2

Min 271.5 F 28.8
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 224.7 F 0.0
10:00 229.1 F 0.0
15:00 230.9 F 0.0
20:00 226.5 F 0.0
25:00 226.7 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 34:46
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:13:25
CYCLE COMPLETE

02/27/2024 10:21 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____