

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003124  
9:34 10/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 0:53  
PRESSURIZING 1:00  
STERILIZING 2:01  
136.4 C 221kPa 2:01  
136.3 C 224kPa 2:31  
136.3 C 223kPa 3:01  
136.2 C 219kPa 3:31  
136.3 C 221kPa 4:01  
136.4 C 223kPa 4:31  
136.1 C 220kPa 5:01  
136.4 C 222kPa 5:31  
136.4 C 222kPa 5:31  
VENTING 5:31  
AIR DRYING 6:23  
CYCLE COMPLETE 6:27

-----  
Digital Signature #  
3310009C3EC00D40  
-----