
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001349
11:47 18/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:39
STERILIZING 3:35
136.4 C 221kPa 3:35
136.1 C 223kPa 4:05
136.5 C 225kPa 4:35
136.4 C 224kPa 5:05
136.4 C 224kPa 5:35
136.4 C 224kPa 6:05
136.4 C 221kPa 6:35
136.4 C 224kPa 7:05
136.4 C 224kPa 7:05
VENTING 7:05
AIR DRYING 7:56
CYCLE COMPLETE 8:00

Digital Signature #
8581328877825B0E
