
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001945
10:21 12/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:15
STERILIZING 4:25
136.0 C 222kPa 4:25
136.4 C 223kPa 4:55
136.3 C 224kPa 5:25
136.2 C 224kPa 5:55
136.4 C 222kPa 6:25
136.4 C 225kPa 6:55
136.3 C 225kPa 7:25
136.4 C 225kPa 7:55
136.4 C 225kPa 7:55
VENTING 7:55
AIR DRYING 8:50
CYCLE COMPLETE 9:15

Digital Signature #
E2722C448867EC20
