
Midmark M11 - v1.0.5
Total Cycles: 3969

Name: RITTER2
Sterilizer ID: V2313228
09/24/2024 08:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 70.7 F 0.0
4:00 75.3 F 0.0
6:00 104.0 F 0.1
8:00 142.6 F 0.4
10:00 173.6 F 0.9
12:00 192.2 F 4.0
14:00 203.2 F 7.5
16:00 217.4 F 11.0
18:00 223.9 F 6.6
20:00 232.6 F 10.0
22:00 235.6 F 9.6
24:00 242.6 F 12.3
26:00 240.8 F 10.9
28:00 253.2 F 17.1
30:00 263.8 F 23.5
32:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.2
2:00 271.4 F 28.6
3:00 272.5 F 29.3
4:00 271.7 F 28.5

Min 271.3 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.4 F 0.0
5:00 198.0 F 0.0
10:00 198.8 F 0.0
15:00 189.7 F 0.0
20:00 186.3 F 0.0
25:00 183.0 F 0.0
30:00 182.7 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:11
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:10:24
CYCLE COMPLETE

09/24/2024 09:43 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____