
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002448
9:12 24/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:37
STERILIZING 3:45
136.0 C 220kPa 3:45
135.9 C 221kPa 4:15
136.3 C 224kPa 4:45
136.2 C 223kPa 5:15
136.4 C 225kPa 5:45
136.3 C 224kPa 6:15
136.3 C 223kPa 6:45
136.4 C 222kPa 7:15
136.4 C 222kPa 7:15
VENTING 7:15
AIR DRYING 8:05
CYCLE COMPLETE 38:05

Digital Signature #
22BD7E98FC75CBB7
