
Midmark M11 - v1.0.5
Total Cycles: 1452

Name: RITTER2
Sterilizer ID: V2313228
09/02/2022 03:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 189.8 F 3.3
4:00 193.5 F 8.1
6:00 219.5 F 13.4
8:00 228.2 F 9.4
10:00 237.4 F 9.3
12:00 243.1 F 13.2
14:00 243.3 F 12.4
16:00 256.2 F 19.2
18:00 267.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.4 F 29.5
2:00 271.3 F 29.1
3:00 272.3 F 29.6
4:00 271.5 F 28.9

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 220.6 F 0.0
10:00 233.3 F 0.0
15:00 234.2 F 0.0
20:00 226.5 F 0.0
25:00 231.3 F 0.0
30:00 218.6 F 0.0

Drying Complete

FILLING: 1:51
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

09/02/2022 04:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____