
Midmark M11 - v1.0.0
Total Cycles: 1665

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2022 05:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.3 F 0.0
2:00 191.0 F 3.2
4:00 195.6 F 7.7
6:00 217.2 F 13.4
8:00 227.9 F 8.0
10:00 240.1 F 13.2
12:00 239.7 F 10.7
14:00 237.3 F 9.0
16:00 251.5 F 15.8
18:00 263.2 F 22.6
20:00 272.1 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.6 F 28.8
2:00 271.6 F 28.0
3:00 272.3 F 28.8
4:00 272.3 F 28.6

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 225.7 F 0.0
10:00 236.9 F 0.0
15:00 221.9 F 0.0
20:00 215.6 F 0.0
25:00 214.8 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 20:10
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

02/24/2022 07:51 AM
ACCEPTED: X
REJECTED:
Operator: U02