

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002400  
10:28 20/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:16  
PRESSURIZING 2:24  
STERILIZING 3:31  
136.4 C 220kPa 3:31  
136.4 C 224kPa 4:01  
136.1 C 223kPa 4:31  
136.4 C 225kPa 5:01  
136.4 C 225kPa 5:31  
136.4 C 225kPa 6:01  
136.4 C 223kPa 6:31  
136.0 C 223kPa 7:01  
136.0 C 223kPa 7:01  
VENTING 7:01  
AIR DRYING 7:51  
CYCLE COMPLETE 9:51  
-----

Digital Signature #  
971CC124BEE4E4E7  
-----