
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000767
9:08 09/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:43
STERILIZING 4:56
136.0 C 220kPa 4:56
136.4 C 223kPa 5:26
136.1 C 217kPa 5:56
136.4 C 224kPa 6:26
136.4 C 223kPa 6:56
136.4 C 223kPa 7:26
136.4 C 225kPa 7:56
136.3 C 224kPa 8:26
136.3 C 224kPa 8:26
VENTING 8:26
AIR DRYING 9:20
CYCLE COMPLETE 9:24

Digital Signature #
467AD698285C0860
