
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001171
13:43 18/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:52
STERILIZING 5:04
136.0 C 220kPa 5:04
136.3 C 224kPa 5:34
136.3 C 224kPa 6:04
136.1 C 221kPa 6:34
136.5 C 225kPa 7:04
136.4 C 225kPa 7:34
136.4 C 223kPa 8:04
136.3 C 223kPa 8:34
136.3 C 223kPa 8:34
VENTING 8:34
AIR DRYING 9:28
CYCLE COMPLETE 24:28

Digital Signature #
CFE0307EDEED72D1
