
Midmark M11 - v1.0.0
Total Cycles: 3169

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2023 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.8 F 0.0
2:00 185.5 F 1.7
4:00 210.9 F 7.1
6:00 223.0 F 12.9
8:00 231.2 F 8.7
10:00 238.0 F 10.0
12:00 245.0 F 12.8
14:00 243.5 F 11.8
16:00 257.1 F 18.9
18:00 268.9 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.2 F 28.5
2:00 272.6 F 28.9
3:00 271.9 F 28.2
4:00 272.8 F 28.9

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 218.5 F 0.0
10:00 228.8 F 0.0
15:00 234.6 F 0.0
20:00 226.1 F 0.0
25:00 224.2 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:08
CYCLE COMPLETE

04/27/2023 01:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____