

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3322

Name: RITTER 1  
Sterilizer ID: V2178198  
06/01/2023 12:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.7 F 0.0  
2:00 194.4 F 4.4  
4:00 206.3 F 9.7  
6:00 226.7 F 6.9  
8:00 232.8 F 10.6  
10:00 233.3 F 7.6  
12:00 242.5 F 11.6  
14:00 245.8 F 13.1  
16:00 258.5 F 19.9  
18:00 270.0 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.4 F 28.9  
2:00 271.5 F 28.1  
3:00 272.5 F 29.1  
4:00 272.1 F 28.6  
  
Min 271.4 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 227.0 F 0.0  
10:00 229.8 F 0.0  
15:00 228.8 F 0.0  
20:00 228.4 F 0.0  
25:00 227.6 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 18:55  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:58:08  
CYCLE COMPLETE

06/01/2023 01:45 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_