

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2486

Name: RITTER 1  
Sterilizer ID: V2178198  
10/25/2022 12:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.7 F 0.0  
2:00 163.0 F 2.2  
4:00 179.8 F 6.1  
6:00 198.0 F 11.1  
8:00 213.6 F 3.8  
10:00 224.6 F 8.6  
12:00 235.8 F 13.2  
14:00 235.7 F 9.8  
16:00 239.6 F 8.8  
18:00 247.0 F 14.1  
20:00 259.2 F 20.7  
22:00 270.2 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.7  
1:00 272.0 F 29.1  
2:00 272.7 F 29.3  
3:00 271.8 F 28.6  
4:00 272.3 F 29.4

Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 202.0 F 0.0  
10:00 205.5 F 0.0  
15:00 209.3 F 0.0  
20:00 199.6 F 0.0  
25:00 195.6 F 0.0  
30:00 191.0 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 22:45  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:01:13  
CYCLE COMPLETE

10/25/2022 01:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_