
Midmark M11 - v1.0.5
Total Cycles: 3634

Name: RITTER2
Sterilizer ID: V2313228
05/22/2024 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.7 F 0.0
2:00 95.7 F 0.1
4:00 106.4 F 1.0
6:00 116.3 F 3.2
8:00 142.4 F 6.2
10:00 166.9 F 9.4
12:00 185.2 F 12.9
14:00 207.6 F 5.1
16:00 220.7 F 8.5
18:00 230.6 F 12.4
20:00 230.6 F 8.8
22:00 241.7 F 13.1
24:00 240.9 F 11.8
26:00 251.7 F 17.2
28:00 261.6 F 22.9
30:00 270.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 272.5 F 30.1
2:00 271.3 F 29.5
3:00 272.4 F 30.2
4:00 271.6 F 29.6

Min 271.2 F 29.5
Max 272.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 204.6 F 0.0
10:00 212.2 F 0.0
15:00 206.0 F 0.0
20:00 208.4 F 0.0
25:00 218.1 F 0.0
30:00 216.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:48
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:09:24
CYCLE COMPLETE

05/22/2024 11:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____