
Midmark M11 - v1.0.5
Total Cycles: 1546

Name: RITTER2
Sterilizer ID: V2313228
09/27/2022 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.5 F 0.0
2:00 168.2 F 3.0
4:00 182.8 F 7.9
6:00 204.8 F 9.5
8:00 229.2 F 9.8
10:00 233.2 F 7.8
12:00 240.2 F 10.8
14:00 246.8 F 14.1
16:00 260.3 F 21.7
18:00 272.2 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.9
1:00 272.9 F 30.3
2:00 272.4 F 30.3
3:00 272.8 F 30.2
4:00 272.3 F 30.2

Min 272.1 F 29.7
Max 273.1 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.7 F 0.0
5:00 219.1 F 0.0
10:00 238.4 F 0.0
15:00 234.7 F 0.0
20:00 226.5 F 0.0
25:00 224.7 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 3:22
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

09/27/2022 04:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____