
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001446
12:12 09/22/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:09
PRESSURIZING 7:22
STERILIZING 9:10
136.0 C 220kPa 9:10
136.4 C 222kPa 9:40
136.3 C 220kPa 10:10
136.4 C 222kPa 10:40
136.3 C 220kPa 11:10
136.1 C 219kPa 11:40
136.4 C 221kPa 12:10
136.4 C 221kPa 12:40
VENTING 12:40
AIR DRYING 13:46
CYCLE COMPLETE 43:46

Digital Signature #
E67F861E9E621F68
