

-----  
Midmark M11 - v1.0.0  
Total Cycles: 797

Name: RITTER 1  
Sterilizer ID: V2178198  
04/22/2021 06:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.5 F 0.0  
2:00 79.9 F 0.0  
4:00 90.5 F 0.0  
6:00 103.8 F 0.0  
8:00 132.7 F 0.1  
10:00 168.4 F 0.2  
12:00 189.5 F 0.3  
14:00 201.3 F 2.7  
16:00 210.6 F 5.1  
18:00 217.3 F 7.7  
20:00 227.0 F 10.8  
22:00 234.3 F 8.5  
24:00 237.8 F 10.6  
26:00 237.8 F 9.0  
28:00 246.5 F 13.5  
30:00 244.3 F 12.2  
32:00 255.2 F 17.9  
34:00 265.1 F 23.8  
36:00 271.7 F 28.0

STERILIZING

mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.3 F 28.7  
2:00 272.4 F 28.7  
3:00 271.8 F 28.2  
4:00 271.4 F 28.1

Min 271.4 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start

mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 218.8 F 0.0  
10:00 219.6 F 0.0  
15:00 207.6 F 0.0  
20:00 227.4 F 0.0  
25:00 231.9 F 0.0  
30:00 227.6 F 0.0

Drying Complete

FILLING: 0:60  
HEATING: 36:09  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:14:47  
CYCLE COMPLETE

04/22/2021 08:02 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_