

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5130

Name: RITTER 1  
Sterilizer ID: V2178198  
11/25/2024 11:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.6 F 0.0  
2:00 70.1 F 0.0  
4:00 85.5 F 0.0  
6:00 122.0 F 0.0  
8:00 158.9 F 0.2  
10:00 186.6 F 0.5  
12:00 207.8 F 3.3  
14:00 216.5 F 6.7  
16:00 226.5 F 10.3  
18:00 231.4 F 8.5  
20:00 238.1 F 10.6  
22:00 234.8 F 7.7  
24:00 242.5 F 10.8  
26:00 247.8 F 13.8  
28:00 260.6 F 20.9  
30:00 272.2 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.9  
1:00 272.6 F 28.5  
2:00 271.6 F 28.0  
3:00 272.6 F 28.5  
4:00 271.5 F 27.9

Min 271.5 F 27.8  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.5 F 0.0  
5:00 191.0 F 0.0  
10:00 187.3 F 0.0  
15:00 184.8 F 0.0  
20:00 183.8 F 0.0  
25:00 187.5 F 0.0  
30:00 200.2 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 30:29  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:08:32  
CYCLE COMPLETE

11/25/2024 12:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_