
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000362
14:33 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:31
STERILIZING 3:37
136.4 C 232kPa 3:37
136.2 C 231kPa 4:07
136.1 C 230kPa 4:37
136.4 C 233kPa 5:07
136.4 C 233kPa 5:37
136.4 C 233kPa 6:07
136.4 C 233kPa 6:37
136.2 C 232kPa 7:07
136.2 C 232kPa 7:07
VENTING 7:08
AIR DRYING 7:59
CYCLE COMPLETE 8:07

Digital Signature #
BD10B88F91A239E6
