
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001037
14:09 13/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:01
STERILIZING 3:08
136.5 C 224kPa 3:08
136.3 C 224kPa 3:38
136.4 C 224kPa 4:08
136.4 C 225kPa 4:38
136.4 C 226kPa 5:08
136.2 C 222kPa 5:38
136.4 C 224kPa 6:08
136.3 C 223kPa 6:38
136.3 C 223kPa 6:38
VENTING 6:38
AIR DRYING 7:29
CYCLE COMPLETE 21:47

Digital Signature #
C1EF2927653073AA
