

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3583

Name: RITTER 1  
Sterilizer ID: V2178198  
07/26/2023 07:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.2 F 0.0  
2:00 73.0 F 0.0  
4:00 80.5 F 0.0  
6:00 113.7 F 0.0  
8:00 149.2 F 0.9  
10:00 175.9 F 3.0  
12:00 189.0 F 5.7  
14:00 206.4 F 8.9  
16:00 217.1 F 12.2  
18:00 222.1 F 5.5  
20:00 230.0 F 9.5  
22:00 240.3 F 11.9  
24:00 239.2 F 10.4  
26:00 236.0 F 8.2  
28:00 247.6 F 13.9  
30:00 258.4 F 19.7  
32:00 268.7 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.2 F 28.5  
2:00 271.8 F 28.6  
3:00 271.9 F 28.3  
4:00 272.0 F 28.8  
  
Min 271.2 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
5:00 206.1 F 0.0  
10:00 214.4 F 0.0  
15:00 225.3 F 0.0  
20:00 227.9 F 0.0  
25:00 225.2 F 0.0  
30:00 225.0 F 0.0  
  
Drying Complete

FILLING: 0:45  
HEATING: 33:11  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:11:07  
CYCLE COMPLETE

07/26/2023 08:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_