
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008024
10:52 05/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:51
STERILIZING 4:58
136.0 C 220kPa 4:58
136.3 C 224kPa 5:28
136.0 C 221kPa 5:58
136.2 C 223kPa 6:28
136.2 C 220kPa 6:58
136.4 C 224kPa 7:28
136.4 C 222kPa 7:58
136.2 C 221kPa 8:28
136.2 C 221kPa 8:28
VENTING 8:28
AIR DRYING 9:30
CYCLE COMPLETE 9:55

Digital Signature #
2AC704EACB7129FA
