

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2368

Name: RITTER 1  
Sterilizer ID: V2178198  
09/26/2022 10:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.2 F 0.0  
2:00 194.8 F 4.0  
4:00 202.7 F 8.8  
6:00 213.7 F 8.5  
8:00 227.4 F 8.3  
10:00 237.6 F 13.1  
12:00 238.9 F 10.8  
14:00 236.7 F 9.0  
16:00 250.6 F 15.9  
18:00 262.5 F 22.7  
20:00 272.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.5 F 29.3  
2:00 271.6 F 28.6  
3:00 272.4 F 29.5  
4:00 272.1 F 29.0  
  
Min 271.5 F 28.6  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 224.7 F 0.0  
10:00 234.3 F 0.0  
15:00 231.8 F 0.0  
20:00 224.8 F 0.0  
25:00 225.8 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 20:17  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:32  
CYCLE COMPLETE

09/26/2022 11:35 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_