
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001458
8:43 06/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:33
STERILIZING 4:33
136.4 C 220kPa 4:33
136.4 C 224kPa 5:03
136.3 C 222kPa 5:33
136.4 C 222kPa 6:03
136.4 C 222kPa 6:33
136.4 C 223kPa 7:03
136.4 C 223kPa 7:33
136.3 C 220kPa 8:03
136.3 C 220kPa 8:03
VENTING 8:04
AIR DRYING 9:00
CYCLE COMPLETE 9:05

Digital Signature #
15EE62F5F53F792B
