
Midmark M11 - v1.0.5
Total Cycles: 3875

Name: RITTER2
Sterilizer ID: V2313228
08/20/2024 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 81.8 F 0.0
4:00 79.5 F 0.0
6:00 116.1 F 0.6
8:00 146.3 F 3.3
10:00 167.8 F 6.6
12:00 216.1 F 10.1
14:00 204.2 F 9.5
16:00 217.4 F 5.4
18:00 225.5 F 8.8
20:00 236.7 F 12.8
22:00 234.3 F 9.3
24:00 240.0 F 10.6
26:00 245.2 F 13.2
28:00 256.5 F 19.3
30:00 266.9 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.7 F 29.3
2:00 271.9 F 29.2
3:00 272.2 F 29.1
4:00 272.2 F 29.4

Min 271.6 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 219.4 F 0.0
10:00 195.6 F 0.0
15:00 186.1 F 0.0
20:00 183.3 F 0.0
25:00 182.0 F 0.0
30:00 180.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 31:33
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:09:53
CYCLE COMPLETE

08/20/2024 09:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____