

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2423

Name: RITTER 1  
Sterilizer ID: V2178198  
10/12/2022 04:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
2:00 185.6 F 4.4  
4:00 211.2 F 9.3  
6:00 224.5 F 6.5  
8:00 232.6 F 9.5  
10:00 238.1 F 9.1  
12:00 242.5 F 12.5  
14:00 241.4 F 11.2  
16:00 254.9 F 18.0  
18:00 266.4 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.1 F 28.8  
2:00 272.6 F 29.3  
3:00 271.5 F 28.6  
4:00 272.2 F 28.9  
  
Min 271.4 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 231.9 F 0.0  
10:00 228.2 F 0.0  
15:00 225.9 F 0.0  
20:00 218.0 F 0.0  
25:00 213.9 F 0.0  
30:00 213.0 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 19:32  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:35  
CYCLE COMPLETE

10/12/2022 05:04 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_