
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000710
9:56 20/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:16
STERILIZING 4:28
136.5 C 232kPa 4:28
136.3 C 232kPa 4:58
136.4 C 233kPa 5:28
136.4 C 233kPa 5:58
136.4 C 233kPa 6:28
136.3 C 233kPa 6:58
136.1 C 230kPa 7:28
136.3 C 232kPa 7:58
136.3 C 232kPa 7:58
VENTING 7:58
AIR DRYING 8:52
CYCLE COMPLETE 8:59

Digital Signature #
F2AEF04742EA0F57
