

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2011

Name: RITTER2  
Sterilizer ID: V2313228  
02/06/2023 01:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.0 F 0.0  
2:00 135.6 F 0.0  
4:00 153.7 F 0.3  
6:00 198.9 F 1.3  
8:00 213.3 F 4.5  
10:00 225.2 F 8.9  
12:00 233.4 F 11.2  
14:00 239.4 F 11.1  
16:00 237.8 F 9.6  
18:00 238.2 F 9.1  
20:00 251.8 F 16.3  
22:00 264.3 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.2 F 28.8  
2:00 272.4 F 29.3  
3:00 271.5 F 28.4  
4:00 272.6 F 29.2

Min 271.5 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
5:00 222.0 F 0.0  
10:00 221.3 F 0.0  
15:00 230.8 F 0.0  
20:00 227.5 F 0.0  
25:00 223.1 F 0.0  
30:00 220.9 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 23:49  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 01:01:40  
CYCLE COMPLETE

02/06/2023 02:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_