

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3350

Name: RITTER 1  
Sterilizer ID: V2178198  
06/07/2023 11:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
2:00 192.1 F 2.8  
4:00 205.9 F 7.8  
6:00 222.7 F 12.3  
8:00 230.1 F 9.3  
10:00 235.5 F 9.4  
12:00 243.6 F 12.8  
14:00 242.1 F 11.6  
16:00 255.4 F 18.3  
18:00 266.7 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.1  
2:00 271.6 F 28.4  
3:00 272.4 F 29.3  
4:00 272.1 F 28.8  
  
Min 271.4 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 220.4 F 0.0  
10:00 235.0 F 0.0  
15:00 234.2 F 0.0  
20:00 235.6 F 0.0  
25:00 222.5 F 0.0  
30:00 211.9 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 19:32  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:46  
CYCLE COMPLETE

06/07/2023 12:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_