
Midmark M11 - v1.0.5
Total Cycles: 4282

Name: RITTER2
Sterilizer ID: V2313228
01/22/2025 06:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.9 F 0.0
2:00 140.4 F 1.1
4:00 145.7 F 3.7
6:00 166.0 F 7.3
8:00 188.3 F 11.3
10:00 196.7 F 3.4
12:00 213.2 F 7.8
14:00 226.4 F 12.1
16:00 229.4 F 8.3
18:00 241.8 F 13.4
20:00 241.0 F 11.8
22:00 252.6 F 17.6
24:00 263.2 F 23.9
26:00 272.8 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.5
1:00 272.9 F 30.2
2:00 272.1 F 29.6
3:00 271.9 F 30.0
4:00 272.8 F 30.1

Min 271.6 F 29.4
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 183.7 F 0.0
10:00 183.1 F 0.0
15:00 181.1 F 0.0
20:00 187.0 F 0.0
25:00 200.3 F 0.0
30:00 210.5 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 26:18
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:05:18
CYCLE COMPLETE

01/22/2025 07:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____