
Midmark M11 - v1.0.5
Total Cycles: 1608

Name: RITTER2
Sterilizer ID: V2313228
10/17/2022 01:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.3 F 0.0
2:00 171.0 F 3.4
4:00 178.9 F 7.2
6:00 198.8 F 12.1
8:00 218.5 F 6.5
10:00 230.8 F 11.7
12:00 232.7 F 8.7
14:00 240.0 F 10.7
16:00 245.7 F 13.6
18:00 257.1 F 20.1
20:00 268.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.7 F 29.9
2:00 271.3 F 29.4
3:00 273.1 F 30.1
4:00 271.8 F 29.5

Min 271.2 F 29.2
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 214.5 F 0.0
10:00 224.9 F 0.0
15:00 231.0 F 0.0
Dry Cycle Aborted

FILLING: 1:27
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:14
DRYING: 19:02
TOTAL CYCLE: 00:48:58

10/17/2022 02:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____