
Midmark M11 - v1.0.5
Total Cycles: 2383

Name: RITTER2
Sterilizer ID: V2313228
05/17/2023 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.8 F 0.0
2:00 189.0 F 4.1
4:00 202.1 F 9.0
6:00 221.5 F 8.2
8:00 232.1 F 10.7
10:00 232.9 F 8.2
12:00 239.2 F 10.6
14:00 247.1 F 14.0
16:00 259.5 F 21.0
18:00 270.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.4 F 29.7
2:00 272.0 F 29.1
3:00 271.2 F 28.9
4:00 272.2 F 29.7

Min 271.2 F 28.7
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.4 F 0.0
5:00 217.9 F 0.0
10:00 242.2 F 0.0
15:00 240.4 F 0.0
20:00 229.2 F 0.0
25:00 227.5 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 18:49
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:00
CYCLE COMPLETE

05/17/2023 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____