
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004433
10:42 26/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 2:59
STERILIZING 3:55
136.0 C 220kPa 3:55
136.4 C 224kPa 4:25
136.4 C 224kPa 4:55
136.4 C 222kPa 5:25
136.4 C 225kPa 5:55
136.4 C 224kPa 6:25
136.4 C 224kPa 6:55
136.4 C 221kPa 7:25
136.4 C 221kPa 7:25
VENTING 7:25
AIR DRYING 8:27
CYCLE COMPLETE 9:28

Digital Signature #
45502DFE1F4DE43C
