

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1614

Name: RITTER2  
Sterilizer ID: V2313228  
10/18/2022 10:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
2:00 193.1 F 3.4  
4:00 207.8 F 8.9  
6:00 226.6 F 8.5  
8:00 232.2 F 10.6  
10:00 235.2 F 8.1  
12:00 240.8 F 11.0  
14:00 246.1 F 14.2  
16:00 259.3 F 21.4  
18:00 270.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.4  
1:00 272.8 F 30.0  
2:00 271.8 F 29.3  
3:00 272.9 F 30.2  
4:00 272.4 F 29.7  
  
Min 271.8 F 29.3  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 200.8 F 0.0  
10:00 209.1 F 0.0  
Dry Cycle Aborted

FILLING: 2:57  
HEATING: 18:48  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 14:53  
TOTAL CYCLE: 00:43:52

10/18/2022 10:52 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_