
Midmark M11 - v1.0.0
Total Cycles: 712

Name: MIDMARK
Sterilizer ID: V2178198
03/01/2021 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.5 F 0.0
2:00 147.5 F 1.4
4:00 157.5 F 5.0
6:00 187.6 F 9.7
8:00 220.0 F 5.1
10:00 223.6 F 9.1
12:00 234.6 F 9.2
14:00 243.1 F 12.4
16:00 241.3 F 11.0
18:00 255.2 F 18.1
20:00 267.0 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 271.8 F 28.9
2:00 271.7 F 28.9
3:00 271.8 F 28.7
4:00 272.2 F 28.8

Min 271.5 F 28.7
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 233.4 F 0.0
10:00 207.7 F 0.0
15:00 208.2 F 0.0
20:00 203.4 F 0.0
Dry Cycle Aborted

FILLING: 0:59
HEATING: 21:55
STERILIZING: 4:00
VENTING: 2:57
DRYING: 22:55
TOTAL CYCLE: 00:52:47

03/01/2021 11:42 AM
ACCEPTED: X
REJECTED:
Operator: U01