

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2293

Name: RITTER 1  
Sterilizer ID: V2178198  
09/06/2022 01:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.5 F 0.0  
2:00 188.8 F 3.3  
4:00 197.3 F 8.5  
6:00 224.0 F 7.6  
8:00 231.7 F 9.3  
10:00 240.6 F 9.4  
12:00 242.5 F 12.2  
14:00 241.0 F 10.8  
16:00 254.7 F 17.7  
18:00 266.3 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.5 F 28.9  
2:00 271.5 F 28.1  
3:00 272.6 F 29.1  
4:00 272.0 F 28.5  
  
Min 271.5 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 216.5 F 0.0  
10:00 224.5 F 0.0  
15:00 230.0 F 0.0  
20:00 228.7 F 0.0  
25:00 228.6 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 2:29  
HEATING: 19:34  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:19  
CYCLE COMPLETE

09/06/2022 02:21 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_